

CALIBRATION CERTIFICATION

We certify that the calibration accuracies listed below are obtained on equipment, and with methods, that can be traced directly to the US National Institute of Standards and Technology.

FLOWRATE READOUT ACCURACY: $\pm 1\%$ Full Scale

PRESSURE TESTED AT: 150 PSIG

METER SERIAL NUMBER: XXXXXXXX

MODEL NUMBER: 62-9

The calibration listed above was performed under the following conditions:

0 to 500 SCFM LANDFILL GAS

40 to 200°F

0 to 15 PSIG

Line Size: 6" SCH 40 (6.065"ID)

STANDARD CONDITIONS

60 DEG F @ 1 ATM

Signature _____ Date August 1, 2013

THERMAL INSTRUMENT COMPANY, INC.

217 Sterner Mill Road, Treose, PA 19053

Phone: 215-355-8400 Fax: 215-355-1789

Web: www.thermalinstrument.com